



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### **Resource Bolting dba Hyltorc Sales and Services**

**45 Corriveau Avenue, St. Albert, AB T8N 5A3**

**Bay # 5, 380 Taiga Nova Crescent, Fort McMurray, AB T9K 0T4**

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

### **ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

### **Mechanical Calibration** *(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President/Operations Manager

*Initial Accreditation Date:*

March 04, 2020

*Issue Date:*

March 04, 2020

*Expiration Date:*

March 31, 2022

*Accreditation No.:*

103809

*Certificate No.:*

L20-126

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjlab.com](http://www.pjlab.com)*



# Certificate of Accreditation: Supplement

## Resource Bolting dba Hytorc Sales and Services

45 Corriveau Avenue, St. Albert, AB T8N 5A3  
380 Taiga Nova Crescent, Fort McMurray, AB T9K 0T4  
Contact Name: Kelly Berge Phone: 780-459-5004

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Indicating Device <sup>F</sup>	1 000 psi to 10 000 psi	0.33 % of Reading	Pressure Transducer TSD 10KPT A.K.O. WI-103, ASTM E2624-17
Torque Indicating Device <sup>F</sup>	40 lbf·ft to 40 000 lbf·ft	0.79 % of Reading	Torque transducer TSD4011, TSD6500-3 AKO WI-101, ASTM E2624-17, ASTM E2428-15a
	500 lbf·ft to 40 000 lbf·ft	0.93 % of Reading	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.